

## Certificate of Analysis

This lot of product meets the accepted performance criteria recommended in ISO 11140-1: 2014 as a Type 1 Process Indicator.

Product Name: Chemical Process Indicator for Monitoring Irradiation Processes

For use in Monitoring: Irradiation (Gamma and E-Beam)

Appearance: Yellow to Red, 25 mm Label with Indicator Ribbon

LOT R220101 REF CRYR-4

Characteristic	Specification	Result
Radiation Cobalt <sub>60</sub> , 10 kGy	Yellow to Red Color Transition	Pass
Date of Product Performance Verification	January 2022	

Storage and Shelf Life				
+15°C-+30°C	Room Temperature (15°– 30°C)	类	Keep away from sunlight	
20%	20% to 80% Relative Humidity	淡	Protect from heat, radioactive sources, and sterilizing agents	
Shelf Life	36 Months from the date of manufacture			

Disposal: Discard as general waste.

	December 27, 2022
Quality Approval	Date

True Indicating, LLC 946 Kane Street, Suite A Toledo, Ohio 43612 www.trueindicating.com info@trueindicating.com FOR LAB/INDUSTRIAL USE ONLY.