TRUE INDICATING PRODUCTS

Certificate of Analysis

This lot of product meets the accepted performance criteria recommended in ISO 11140-1: 2014 as a Type 1 Process Indicator.

Product Name: Indicator for Monitoring Vaporized Hydrogen Peroxide (VH_2O_2) Processes

LOT P180813 REF CPRY-1T 13-AUG-2021

13-AUG-2018 For use in Monitoring: Vaporized Hydrogen Peroxide

Appearance: **Indicator Tape, 19 mm x 55 m**

Performance Characteristics

Characteristic	Specification	Result
VH_2O_2 50°C, 6 minutes, and 2.3 mg/L	Red to Yellow Color Transition	Pass
Date of Product Performance Verification:	August 2018	

Storage and Shelf Life

+15°C +30°C	Room Temperature (15° – 30°C)		Keep away from sunlight
20% 80%	20% to 80% Relative Humidity		Protect from heat, radioactive sources, and sterilizing agents
Shelf Life	36 Months from the date of manufacture		

Disposal: Discard as general waste.

_____ _____

Quality Approval Date

True Indicating, LLC
946 Kane Street, Suite A
Toledo, Ohio 43612

info@trueindicating.com

www.trueindicating.com